

For rapid qualitative detection of Leukocytes, Blood, Nitrite and Protein in human urine.

For self-testing *in vitro* diagnostic use only.

【INTENDED USE】

The Urinary Tract Infections Test (Urine) is a plastic strips onto which several separate reagent areas are affixed. The test is for the qualitative detection of the following analytes in urine: Blood, Protein, Nitrite and Leukocytes. The Urinary Tract Infections Test (Urine) is for single use in self-testing.

【SUMMARY】

A urinary infection represents the most common disease of the urinary tract which includes the urethra, the bladder, the ureter and the kidneys. Although men, women and children can experience a urinary infection, it's mostly women who suffer from these infections, since the short urethra favours the penetration of germs. Elderly males are also affected if they have an enlarged prostate which obstructs the urine flow.

In healthy people, urine is sterile (i.e. it doesn't contain any micro-organisms). One of the best ways to keep your urinary tract sterile is to empty your bladder completely at regular intervals. Generally, an infection starts in the urethra and may then spread into the upper urinary tract as far as the kidneys.

The symptoms vary considerably: burning when emptying the bladder, or a strong urge to urinate. The urine may also be cloudy or have a strong odour.

【PRINCIPLES OF THE EXAMINATION METHOD】

Leukocytes: This test reveals the presence of granulocyte esterases. The esterases cleave a derivatized pyrazole amino acid ester to liberate derivatized hydroxyl pyrazole. This pyrazole then reacts with a diazonium salt to produce a beige-pink to purple colour.

Blood: This test is based on the peroxidase-like activity of hemoglobin which catalyzes the reaction of diisopropylbenzene dihydroperoxide and 3,3',5,5'-tetramethylbenzidine. The resulting colour from orange to green to dark blue.

Nitrite: This test depends upon the conversion of nitrate to nitrite by the action of Gram negative bacteria in the urine. In an acidic medium, nitrite in the urine reacts with p-arsanilic acid to form a diazonium compound. The diazonium compound in turn couples with 1 N-(1-naphthyl) ethylenediamine to produce a pink colour.

Protein: This reaction is based on the phenomenon known as the "protein error" of pH indicators (Tetrabromopheno Blue). The anion produced by pH indicators under given conditions combines with the cation produced by protein, then pH indicators colours from yellow to green-blue for positive results.

【PRECAUTIONS】
Read the instructions carefully before performing the test.

- For self-test *in vitro* diagnostic use only.
- Store in a dry place at 2-30°C (36-86°F), avoiding areas of excess moisture. If the foil packaging is damaged or has been opened, please do not use.
- A clean container - uncontaminated by cleaning fluids - to collect urine.
- Keep out of the reach of children.
- Do not use after the expiry date or if the pouch is damaged.
- Follow the indicated time strictly.
- Use the test only once. Do not dismantle and touch the reagent areas of the test strip.
- For external use only.
- The used test should be discarded according to local regulations.
- In case of difficulties in colour identification (such as Daltonism), ask for help when interpreting the test.

【STORAGE AND STABILITY】

Store as packaged at room temperature or refrigerated (2-30°C). The test is stable through the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until its use. **DO NOT FREEZE.** Do not use beyond the expiration date.

【MATERIALS PROVIDED】

- Test strip
- Plastic cup
- Colour chart
- Package insert

【MATERIALS REQUIRED BUT NOT PROVIDED】

- Timer or a watch
- Specimen container

【PROCEDURE】

ATTENTION: It's recommended to take the sample from the first urine of the day since it's the most concentrated. The urine used for the test should not come into contact with water from the toilet or any disinfectant or cleaning substances.

For women only: The test should not be performed during or for three days after your menstrual period. The urine sample should not be contaminated with vaginal fluids since this may produce a misleading result.





Do not make any important medical decision without first consulting with your doctor.

COLLECT URINE:

Collect part of the urine in the supplied plastic cup or using a clean cup without any residual detergents. Make sure to fill the cup with urine.

PERFORMING THE TEST:

- 1) Open the foil pouch and take out the test strip. **Do not touch the test fields.** Once the pouch has been opened, it is recommended to perform the test immediately.
- 2) Dip the test strip in the urine sample.
- ATTENTION:** Press the strip and make sure that all four test fields are immersed for about **1-2 seconds**.
- 3) Then remove the test strip and wipe off any surplus urine against the rim of the container or with some absorbent material (e.g. a paper towel) to avoid mixing chemicals from adjacent reagent areas.
- 4) **Wait for 2 min** (do not read results after 3 minutes).



【READING THE RESULTS】

Read the result separately for each parameter; compare colour with colour chart provided.

Colour changes on the edges of the test pads or colour changes after more than 3 minutes should be disregarded.

NEGATIVE

- The Test field for **LEUKOCYTES** stayed whitish.
- The Test field for **BLOOD** stayed mustard yellow
- The Test field for **NITRITE** stayed white.
- The Test field for **PROTEIN** stayed yellowish.

LEUKOCYTES NITRITE

BLOOD PROTEIN
POSITIVE RESULT FOR LEUKOCYTES

If the colour of the test field has changed to **purple**, then leukocytes have been found in your urine.

LEUKOCYTES

POSITIVE RESULT FOR BLOOD

If the colour of the test field has changed to **green** (or some green spots appear on the background), then blood has been found in your urine.

BLOOD

POSITIVE RESULT FOR NITRITE

If the colour of the test field has changed to **pink**, then nitrites have been found in your urine.

NITRITE

POSITIVE RESULT FOR PROTEIN

If the colour of the test field has changed to **green**, then proteins have been found in your urine.

PROTEIN


【TECHNICAL NOTES ON PARAMETERS】

The Test detects **LEUKOCYTES, BLOOD, NITRITE** and/or **PROTEIN** in urine.

LEUKOCYTES: The presence of leukocytes in urine is a symptom of an inflammation of the kidneys and the urinary tract, through a series of chemical reactions the presence of granulocyte esterases change the colour of the pad to purple.

Taking cephalaxin and cephalothin, or high concentration of oxalic acid may cause test results to be artificially low. Tetracycline may cause decreased reactivity, and high levels of the drug may cause a false negative reaction.

High urinary protein may diminish the intensity of the reaction colour.

BLOOD: A uniform green colour conversion indicates the presence of hemoglobin or hemolyzed erythrocytes; scattered or compacted green spots indicate intact erythrocytes. Generally, urine occult blood can be attributed to one of the following three reasons: the presence of stones, inflammation, and/or cancer. Inflammation from glomerulonephritis, pyelonephritis, cystitis, but also hematuria can result in urine occult blood. Kidney stones, ureter or bladder stones, may also result in the presence of occult blood. The presence of tumours can also cause occult blood (either a benign or malignant tumour/s of the kidney, ureter and bladder). Menstrual period, constipation can also lead to a positive result.

NITRITE: Gram-negative bacteria in urine convert nitrate from food into nitrite. Nitrite reacts with a chemical in the test field and leaves a pink shade. The test result may be distorted if urine does not stay long in the bladder, due to hunger, a vegetable-free diet or antibiotic treatment. Comparing the test on a white background may aid in the detection of low nitrite levels, which might otherwise be missed.

PROTEIN: An indicator on the test field reacts with protein in the urine, changing its colour to green. Protein may be found when there is inflammation of the bladder or prostate, bleeding in the urinary tract. Infusions containing polyvinylpyrrolidone may yield a false positive result.

Chemical components in the test fields must be viewed as potentially dangerous substances, although they present no hazard provided that all test components are used in accordance with these instructions.

【CONTROL PROCEDURE】

To serve as good quality control, the instruction must be followed closely when performing the test. Failure to follow provided instructions may yield inaccurate test results.

【PERFORMANCE CHARACTERISTICS】

Parameters of importance to the user are sensitivity, specificity, accuracy and precision. This test has been developed to be specific for the parameters to be measured with the exceptions of the interferences listed. Please refer to the Limitations section in this package insert. Interpretation of visual results is dependent on several factors: the variability of colour perception, the presence or absence of inhibitory factors, and the lighting conditions when the strip is read. Each colour block on the chart corresponds to a range of analyte concentrations.

【LIMITATIONS】

Note: The Urinary tract infections test (Urine) may be affected by substances that cause abnormal urine colour such as drugs containing azo dyes (e.g. Pyridium[®], AzoGantistrin[®], AzoGantano[®]), nitrofurantoin (Microdantin[®], Furdantin[®]), and riboflavin.¹ The colour development on the test pad may be masked or a colour reaction may be produced that could be interpreted as a false result.

Leukocytes: The result should be read at 2 min to allow for complete colour development. The intensity of the colour that develops is proportional to the number of leukocytes present in the urine specimen. High specific gravity or elevated glucose concentrations (≥ 2,000 mg/dL) may cause test results to be artificially low. The presence of cephalaxin, cephalothin, or high concentrations of oxalic acid may also cause test results to be artificially low. Tetracycline may cause decreased reactivity, and high levels of the drug may cause a false negative reaction. High urinary protein may diminish the intensity of the reaction colour. This test will not react with erythrocytes or bacteria common in urine.¹

Blood: A uniform green colour indicates the presence of myoglobin, hemoglobin or hemolyzed erythrocytes.¹ Scattered or compacted green spots indicate intact erythrocytes. To enhance accuracy, separate colour scales are provided for hemoglobin and for erythrocytes. Positive results with this test are often seen with urine from menstruating females. It has been reported that urine of high pH reduces sensitivity, while moderate to high concentration of ascorbic acid may inhibit colour formation. Microbial peroxidase, associated with urinary tract infection, may cause a false positive reaction.² The test is slightly more sensitive to free hemoglobin and myoglobin than to intact erythrocytes.

Nitrite: The test is specific for nitrite and will not react with any other substance normally excreted in urine. Any degree of uniform pink to red colour should be interpreted as a positive result, suggesting the presence of nitrite. Colour intensity is not proportional to the number of bacteria present in the urine specimen. Pink spots or pink edges should not be interpreted as a positive result. Comparing the reacted reagent area on a white background may aid in the detection of low nitrite levels, which might otherwise be missed. Ascorbic acid above 30 mg/dL may cause false negatives in urine containing less than 0.05 mg/dL nitrite ions. The sensitivity of this test is reduced for urine specimens with highly buffered alkaline urine or with high specific gravity. A negative result does not at any time preclude the possibility of bacteriuria. Negative results may occur in urinary tract infections from organisms that do not contain reductase to convert nitrate to nitrite; when urine has not been retained in the bladder for a sufficient length of time (at least 4 hours) for reduction of nitrate to nitrite to occur; when receiving antibiotic therapy or when dietary nitrate is absent.³

Protein: Any green colour indicates the presence of protein in the urine. This test is highly sensitive for albumin, and less sensitive to hemoglobin, globulin and mucoprotein.¹ A negative result does not rule out the presence of these other proteins.

False positive results may be obtained with highly buffered or alkaline urine. Contamination of urine specimens with quaternary ammonium compounds or skin cleansers containing chlorhexidine may produce false positive results.¹ Urine samples with high specific gravity may provide false negative results.

【EXTRA INFORMATION】

WHAT SHOULD I DO IF MY TEST RESULT IS POSITIVE?





Remember that a positive result doesn't mean that all four substances have been detected in your urine. Even if your result is positive for just one of them, it is important to consult with a medical professional. Get in touch promptly with your own doctor, who will be able to provide a more accurate diagnosis. When you visit your doctor, please take these instructions with you so that they will be better informed as to the type of test you have performed.

WHAT SHOULD I DO IF MY TEST RESULT IS NEGATIVE?





Remember that your test result is only negative if the result on the test field for all four substances are negative. But if continue to experience the signs/symptoms of a UTI or have any other symptoms, then contact a doctor/medical professional to arrange a more thorough examination.




【BIBLIOGRAPHY】

1. Henry JB, et al. Clinical Diagnosis and Management by Laboratory Methods, 20th Ed. Philadelphia. Saunders. 371-372, 375, 379, 382, 385, 2001.
2. Ma Junlong, Cong Yulong. The effect of bacteriuria on the determination of urine red blood cells by urine analyzer. Chinese Journal of Medical Examination, 1999, 22(4): 205.
3. Shuai Lihua, Jiujiang Medical Journal 2002, 17 (2): 122.

	Consult Instructions for Use
	For <i>in vitro</i> diagnostic use only
	Store between 2-30°C
	Do not use if package is damaged

Index of Symbols

	Tests per kit
	Use by
	Lot Number
	Manufacturer

	Authorized Representative
	Do not reuse
	Catalog #



Hangzhou AllTest Biotech Co.,Ltd.
 #550, Yinhai Street
 Hangzhou Economic & Technological Development Area
 Hangzhou, 310018 P.R. China
 Web: www.alltests.com.cn Email: info@alltests.com.cn



MedNet EC-REP GmbH
 Borkstrasse 10,
 48163 Muenster,
 Germany

Number: 146804600
 Effective date: